

Air quality monitoring for volcanic emissions: briefing document

Volcanic ash and gas are types of airborne pollution and should be monitored to check if air quality is exceeding public health standards. Government agencies may want to use air quality data to provide public information and recommend actions associated with different levels of an air quality index (e.g. <u>https://www.airnow.gov/aqi/aqi-basics/</u>).

Air quality monitors may be real-time or may provide retrospective data. Real-time monitors are more useful for immediate public health advisories.

Ideally, regulatory-grade instruments would be installed but these are expensive, challenging and time-consuming to install in a crisis, and need regular maintenance. Cost also prohibits installation of a monitoring network. Low-cost sensors are a good alternative during a crisis and are suitable for establishing a network, with sensors across populated centres. The data are not as reliable as regulatory data but may be used as an indication of air quality, rather than absolute values. Therefore, advisories associated with the data should be given with caution. The monitors recommended below are all low-to-medium cost.

Particulate monitoring

	DustTrak	PurpleAir	
	DustTrak DRX Aerosol Monitor 8533		
Cost	Approx. \$U\$ 15.000	ŚUS 279 per unit	
What they measure/report	Simultaneous measurement of PM ₁ , PM _{2.5} , PM ₄ , PM ₁₀ , TSP	Simultaneous measurement of PM_1 , $PM_{2.5}$, PM_{10} , Temperature, Humidity Derived air quality indices (e.g., USEPA $PM_{2.5}$ AQI)	
Power requirements	Battery life is approximately 6-8 hours.	Work best with mains power. Can be run for several days from a power bank. Can be run from vehicle with USB outlet.	
Data access	Data must be downloaded from instrument.	With access to wifi, data can be uploaded to PurpleAir's cloud where it can be visualised in real time, on a	

The two main low-to-medium cost **non-regulatory** instruments used for airborne particulate monitoring during volcanic eruptions are the TSI DustTrakaerosol monitor and PurpleAir sensors.



		 map with different averaging periods and with clear links to health information and advice. Without wifi, the SD card-enabled devices save data to a microSD card which can be extracted and the data downloaded.
Are they regulatory quality?	No, but the DustTrak is used extensively for air quality research around the world, including during volcanic eruption (e.g. the Soufrière Hills).	No, but the PurpleAir has been <u>tested</u> against other low-cost PM sensors and performs well and is an excellent choice for setting up a network of sensors.
Ease of use and maintenance	Easy to use, very little training required. Requires daily calibration and data download every few days. Should be placed in a sheltered (from rainfall/ash) environment such as a veranda. Environmental enclosures can be purchased.	Easy to use, very little training required. Requires data download every few days if no wifi. Installation is straightforward but important to consider location. Should be checked and cleaned (to remove insects) frequently.

Gas monitoring

The following low-cost gas detection techniques are **non-regulatory** methods for assessing gas concentrations. The methods presented below can help characterize gas concentrations over time and space, for the purposes of evaluating hazards. These instruments have a higher minimum detection limit and lower resolution than regulatory methods but are cheaper and require less infrastructure.

Technique	Gas detection tubes	Personal or handheld gas monitors	Better resolution or broadcast capable sensors
		Dock station	AreaRAE Aeroqual
Cost	~US \$500 for Hand pump + \$100-150/10 tubes	~US \$400-1,000/ monitor (+ ~\$1,000 – \$3,000 for docking station/calibration gases, except Aeroqual)	~US \$10,000 – \$20,000



Example	Drager, RAE systems,	Industrial scientific, BW	RAE systems (<u>areaRAE</u> or
manufacturers/products	<u>Sensidyne</u>	technologies, Drager,	<u>multiRAE</u>), Interscan
		Aeroqual	corporation, (<u>GasD 8240</u>),
			<u>Aeroqual</u>
What they measure	Single gas concentration in	Single or multiple gas	Single or multiple gas
	ppm (i.e SO ₂ , H_2S)	concentration in ppm	concentration in ppm
		(i.e. SO_2 , H_2S , CO_2)	$(SO_2, H_2S, CO_2, others)$
Resolution	0.1 - 1 ppm	0.1 ppm	0.1 - 0.01 ppm
Range	0.1 - 200 ppm options	Generally, 0 - 200 ppm	SO ₂ : 0-2, 20, or 100 ppm
Power requirement	none	Internalbattery	AC, limited internal
			battery, or configure for
			external battery
Data access	Direct readout on tube	Screen readout +	Log onsite or telemetered
		logging capability. Data	
		download via stand-	
		alone or networked	
		docking station	
How to use	Operator needed for manual	Can use for personal	Deploy to field location for
	pumping of gas through tube.	protection or field	time series data. Some
	3-20 min/sample	deploy in environmental	have wireless radio option
		enclosure	for real-time data
			acquisition
Ease of use and	Easy, little training required.	Easy. Basic technical skill	Basic technical expertise
maintenance	Used tubes may be	needed for docking	required. Aeroqual
	considered chemical waste so	station set-up and	monthly/quarterly
	consult local regulations for	badge programming.	maintenance: filter
	disposal.	Monthly calibration	replacement, flow/leak
		recommended.	and calibration checks,
		Calibration gas cylinders	clean inlet. Pump
		required.	replacement 12-18
			months. AreaRAE
			maintenance as needed:
			sensor, filter, battery,
			pump replacement.
Limitations	Single data point manually	Not wifi enabled -	Wireless capability is 2-3
	collected in time and space.	manual data download	km. Some configurable
		required, except	with 900 MHz radio or
		Aeroqual 300/500 has	cloud-based user interface
		voltage output/relays	for real-time, remote
		for datalogger interface.	access.

Written by Carol Stewart (Massey University, NZ), Tamar Elias (US Geological Survey), Claire J. Horwell (Durham University, UK), David Damby (US Geological Survey). Last edited 15 April 2021. Version 1.5